



TEST REPORT TUCHENG

Date: Jul. 31, 2020 Date of Receipt: Jun. 17, 2020

Report No.: TAF9F465-B Quantity: 1PC Page Order/Pages: (P1/1) Ref. No.: NIL


Report Title: MAGICAL FILM ENTERPRISE CO., LTD Item: Filter Media

Address: No. 36, Aly. 378, Sec. 2, Yongxing Rd., Wuqi Dist., Taichung City 435, Taiwan (R.O.C.)

Test Items		Test Results	Test Methods
Sub-Micron Particulate Filtration Efficiency(%) (0.1 μm PSL)	1	94.08	ASTM F2100-2019 9.3
	2	94.37	ASTM F2299-2017
	3	94.29	Flow rate:28.1
	4	94.17	(Liter/min)
	5	94.11	
	Ave.	94.20	

Note: Sample description is given by the client: Meltblown NO.K002

Note: The test report is the test results issued by the testing institution as requested by the consignor, it shall not determine the legitimacy of the product.



- Note: 1.This report is only responsible for the submitted sample(s), which will be kept for one month period.  
2.This report cannot be reproduced in any way, except in full context, without the prior approval in writing of this Department of Testing and Certification.  
3.The test report should not be used for public advertisement and commercial promotion.

Authorized by president of  
Taiwan Textile Research Institute

**Jui-hung Kao**  
Director,  
Department of Testing and  
Certification

Department of Testing and Certification, Taiwan Textile Research Institute  
No.6, Chengtian Rd., Tucheng Dist., New Taipei City 23674, Taiwan (R.O.C.)  
Tel : +886-2-22670321 ext. 7107, 7110  
Fax : +886-2-22675108 , +886-2-22689839