



正本  
ORIGINAL

日期 2020.09.07 收件日期 2020.08.24 試驗報告 土城場區  
Date: Date of Receipt: TEST REPORT TUCHENG


報告編號 TFF9H637 數量 1件 報告頁次/頁數 (P1/1) 來文字號 空白  
Report No.: Quantity: Page Order/Pages: Ref. No.:

報告抬頭 彩麗薄膜股份有限公司(U2949) 試件類別 濾材  
Report Title: Item:

地址 435 台中市梧棲區永興路二段378巷36號  
Address:

試驗項目	試驗結果	試驗方法
微粒過濾效率 (0.1 μm PSL)	1	98.19
	2	98.26
	3	98.26
	4	98.25
	5	98.23
	平均	98.24
		ASTM F2100-2019 9.3
		ASTM F2299-2017
		流量:28.1 (Liter/min)

註：依委託者所提供來樣資料為：熔噴布 056-5



附記: 1.本報告僅對樣品負責，樣品保留期限為一個月。  
Note: This report is only responsible for the submitted sample(s), which will be kept for one month period.

2.本報告非經本檢測及驗證部書面同意，不得摘錄複製，但全部複製除外。  
This report cannot be reproduced in any way, except in full context, without the prior approval in writing of this Department of Testing and Certification.

3.試驗報告所載事項不得作為公開廣告及商業推銷之用。  
The test report should not be used for public advertisement and commercial promotion.

財團法人紡織產業綜合研究所  
所長授權核發人:  
Authorized by president of  
Taiwan Textile Research Institute

紡織產業綜合研究所檢測及驗證部 新北市土城區承天路6號  
Department of Testing and Certification, Taiwan Textile Research Institute  
No.6, Chengtian Rd., Tucheng Dist., New Taipei City 23674, Taiwan (R.O.C.)  
Tel : +886-2-22670321 ext. 7107, 7110  
Fax : +886-2-22675108, +886-2-22689839



正本  
ORIGINAL


日期 2020.09.07 收件日期 2020.08.24  
Date: Date of Receipt:

試驗報告 土城場區  
TEST REPORT TUCHENG

報告編號 TFF9H638 數量 1件 報告頁次/頁數 (P1/1) 來文字號 空白  
Report No.: Quantity: Page Order/Pages: Ref. No.:  
報告抬頭 彩麗薄膜股份有限公司(U2949) 試件類別 濾材  
Report Title: Item:  
地址 435 台中市梧棲區永興路二段378巷36號  
Address:

試驗項目	試驗結果	試驗方法
微粒過濾效率 (0.1 μm PSL)	1	97.90
	2	97.85
	3	97.87
	4	97.89
	5	97.92
	平均	97.88
		ASTM F2100-2019 9.3 ASTM F2299-2017 流量:28.1 (Liter/min)

註：依委託者所提供來樣資料為：熔噴布 056-5 100°C 24h



附記：1.本報告僅對樣品負責，樣品保留期限為一個月。  
Note: This report is only responsible for the submitted sample(s), which will be kept for one month period.  
2.本報告非經本檢測及驗證部書面同意，不得摘錄複製，但全部複製除外。  
This report cannot be reproduced in any way, except in full context, without the prior approval in writing of this Department of Testing and Certification.  
3.試驗報告所載事項不得作為公開廣告及商業推銷之用。  
The test report should not be used for public advertisement and commercial promotion.



財團法人紡織產業綜合研究所  
所長授權核發人：  
Authorized by president of  
Taiwan Textile Research Institute



紡織產業綜合研究所檢測及驗證部 新北市土城區承天路6號  
Department of Testing and Certification, Taiwan Textile Research Institute  
No.6, Chengtian Rd., Tucheng Dist., New Taipei City 23674, Taiwan (R.O.C.)  
Tel : +886-2-22670321 ext. 7107, 7110  
Fax : +886-2-22675108 , +886-2-22689839