



Date: Aug. 17, 2020 Date of Receipt: Jul. 30, 2020 TEST REPORT TUCHENG ORIGINAL

Report No.: TFF9G820 Quantity: 1PC (P1/1) Ref. No.: NIL


Report Title: MAGICAL FILM ENTERPRISE CO., LTD Item: Filter Media

Address: No. 36, Aly. 378, Sec. 2, Yongxing Rd., Wuqi Dist., Taichung City 435, Taiwan (R.O.C.)

Test Items		Test Results	Test Methods
Sub-Micron Particulate Filtration Efficiency (%) (0.1 μm PSL)	1	98.17	ASTM F2100-2019 9.3
	2	98.16	ASTM F2299-2017
	3	97.87	Flow rate:28
	4	98.03	(Liter/min)
	5	97.99	
	Ave.	98.04	

Note: Sample description is given by the client: Meltblown NO.K002

Note: The test report is the test results issued by the testing institution as requested by the consignor, it shall not determine the legitimacy of the product.



- Note: 1.This report is only responsible for the submitted sample(s), which will be kept for one month period.
- 2.This report cannot be reproduced in any way, except in full context, without the prior approval in writing of this Department of Testing and Certification.
- 3.The test report should not be used for public advertisement and commercial promotion.

Authorized by president of Jui-hung Kao  
Taiwan Textile Research Institute

Director,  
Department of Testing and Certification, Taiwan Textile Research Institute  
No.6, Chengtian Rd., Tucheng Dist., New Taipei City 23674, Taiwan (R.O.C.)  
Tel : +886-2-22670321 ext. 7107, 7110  
Fax : +886-2-22675108 , +886-2-22689839